



ISO 9001 Certified

TEST CERTIFICATE

for the

MODEL NUMBER

HQ30d

SERIAL NUMBER 091100036417	DATE TESTED 11/24/2009
--------------------------------------	----------------------------------

	Minimum Limit	Maximum Limit	Actual
KEYPAD TEST			PASS
DISPLAY TEST			PASS
PROBE RECOGNITION			PASS
BATTERY ON CURRENT	0.001 A	0.12 A	0.0102
BATTERY OFF CURRENT	0.000 A	.0002 A	3.0334E-5

Test equipment used for the verification of Hach manufactured instruments is calibrated using standards traceable to National Institute of Standards and Technology (NIST) standards. Where such standards do not exist, the basis for calibration is documented.

Larry Brew-Quality Manager

Steve Kemper-Test Engineer

TEST CERTIFICATE, part of document 11226-20-HQ0001

FOR TECHNICAL ASSISTANCE, PRICE INFORMATION AND ORDERING:

In the U.S.A. - Call toll-free 800-227-4224

Outside the U.S.A. - Contact the HACH office or distributor serving you.

On the Worldwide Web - www.hach.com; E-mail techhelp@hach.com

**HACH COMPANY
WORLD**

HEADQUARTERS

Telephone: (970) 669-3050

FAX: (970) 669-2932