



ISO 9001 Certified  
**TEST CERTIFICATE**  
 For the Conductivity Probe

Model	Serial Number	Date	Quality Check
CDC40101	182132587068	8/1/2018	PASS

Tested Characteristic	Min	Max	Value	
Probe Recognition				PASS
Physical Inspection				PASS
Reference Temperature (°C)	15	30	20.96	PASS
Diff. Temperature probe vs ref. (°C)	-0.3	0.3	0.27	PASS
Calibration Temperature (°C)	15	35	21.23	PASS
Cell Constant (cm-1)	0.37	0.44	0.39	PASS

	Nominal	Type
Standard 1	1000 uS @25°C	NaCl

Test equipment used for the verification of Hach manufactured instruments is calibrated using standards traceable to National Institute of Standards and Technology (NIST) standards. Where such standards do not exist, the basis for calibration is documented.

**FOR TECHNICAL ASSISTANCE, PRICE INFORMATION AND ORDERING:**

In the U.S.A. – Call toll-free 800-227-4224

Outside the U.S.A. – Contact the HACH office or distributor serving you.

On the Worldwide Web – [www.hach.com](http://www.hach.com); E-mail [techhelp@hach.com](mailto:techhelp@hach.com)

**HACH COMPANY**

**WORLD HEADQUARTERS**

Telephone: (970) 669-3050

FAX: (970) 669-2932