

Certificate of Analysis List

For request number 1123713

| Catalog Number Entered | Lot Number Entered | Related Catalog Number | Related Lot Code | Description |
|---------------------------------------|-----------------------------------|---------------------------------------|---------------------------------|--------------------------------------|
| 246142-LM | 9050 | N/A | N/A | Formazin Turbidity Standard 4000 NTU |

Total Enclosures: 1

Confidencial
GRUENTEC CÍA. LTDA.



An ISO 9001 Certified Company

Certificate of Analysis

Page 1

COMMODITY: **Formazin Turbidity Standard 4000 NTU**
 COMMODITY NUMBER: **246142-LM** MANUFACTURE DATE:
 LOT NUMBER: **A9050** **3/4/2019**

DATE OF ANALYSIS:
3/4/2019

| TEST | SPECIFICATIONS | RESULTS |
|---|-----------------------|----------------|
| Turbidity of 40 NTU formazin dilution of this lot compared to lab standard and measured on a Hach 2100 AN Turbidimeter. | 39.2 to 40.8 NTU | 40.60 NTU |
| Turbidity of 100 NTU formazin dilution of this lot compared to lab standard and measured on a Hach 2100 AN Turbidimeter. | 98 to 102 NTU | 102.0 NTU |
| Turbidity of 400 NTU formazin dilution of this lot compared to lab standard and measured on a Hach 2100 AN Turbidimeter. | 392 to 408 NTU | 403.0 NTU |
| Turbidity of 1000 NTU formazin dilution of this lot compared to lab standard and measured on a Hach 2100 AN Turbidimeter. | 980 to 1020 NTU | 1016.0 NTU |
| Turbidity of the standard undiluted. | 3920 to 4080 NTU | 4024.0 NTU |

The expiration date is Feb 2021

Certified by _____

Scott Als
 Analytical Services Chemist