

Certificate of Analysis List

For request number 1270582

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
2974349	0003	N/A	N/A	Potassium Chloride Standard Solution

Total Enclosures: 1

Confidencial
GRUENTEC CÍA. LTDA.



An ISO 9001 Certified Company

Certificate of Analysis

Page 1

COMMODITY: **Potassium Chloride Standard Solution**COMMODITY NUMBER: **2974349**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A0003****1/10/2020****1/10/2020**

TEST	SPECIFICATIONS	RESULTS
conductivity	1398 to 1426 uS/cm	1407.0 uS/cm

The expiration date is Jan 2021

The item 2974349 is traceable to NIST standards SRM 999C Potassium Chloride LOT N/A.

Confidencial
GRUENTEC CÍA. LTDA.

A handwritten signature in cursive script that reads "Scott Als".

Certified by _____

Scott Als
Analytical Services Chemist